
Midmark M11 - v1.0.0
Total Cycles: 3191

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2023 07:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.1 F 0.0
2:00 78.8 F 0.0
4:00 85.4 F 0.0
6:00 118.6 F 0.0
8:00 156.2 F 1.3
10:00 179.5 F 3.5
12:00 192.4 F 6.5
14:00 206.1 F 9.9
16:00 220.3 F 12.3
18:00 226.2 F 7.6
20:00 234.8 F 11.9
22:00 232.4 F 8.3
24:00 243.6 F 12.3
26:00 243.0 F 12.0
28:00 254.6 F 18.0
30:00 265.2 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.6 F 29.3
2:00 271.8 F 28.6
3:00 271.6 F 28.9
4:00 272.6 F 29.2

Min 271.1 F 28.3
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.4 F 0.0
5:00 198.0 F 0.0
10:00 216.5 F 0.0
15:00 198.2 F 0.0
20:00 223.5 F 0.0
25:00 232.8 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:56
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:57
CYCLE COMPLETE

05/03/2023 08:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____