

-----  
Midmark M11 - v1.0.5  
Total Cycles: 138

Name: RITTER2  
Sterilizer ID: V2313228  
06/24/2021 11:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.8 F 0.0  
2:00 143.3 F 1.9  
4:00 151.3 F 4.6  
6:00 178.2 F 8.1  
8:00 197.7 F 12.4  
10:00 213.2 F 6.0  
12:00 226.8 F 10.6  
14:00 229.4 F 8.0  
16:00 239.9 F 12.7  
18:00 240.5 F 11.6  
20:00 252.6 F 17.8  
22:00 263.6 F 24.3  
24:00 272.4 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.7  
1:00 272.8 F 30.4  
2:00 271.9 F 29.7  
3:00 272.3 F 30.5  
4:00 272.5 F 30.2

Min 271.6 F 29.6  
Max 272.9 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
5:00 195.2 F 0.0  
10:00 200.5 F 0.0  
15:00 203.2 F 0.0  
20:00 203.1 F 0.0  
25:00 206.3 F 0.0  
30:00 208.0 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 24:12  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:02:43  
CYCLE COMPLETE

06/24/2021 12:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_