
Midmark M11 - v1.0.0
Total Cycles: 2562

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2022 02:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.0 F 0.0
2:00 194.7 F 4.0
4:00 199.0 F 9.8
6:00 225.2 F 5.3
8:00 235.2 F 10.9
10:00 235.5 F 8.1
12:00 242.6 F 10.5
14:00 247.2 F 13.8
16:00 260.1 F 21.0
18:00 271.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.7 F 29.0
2:00 271.8 F 28.5
3:00 272.8 F 29.1
4:00 272.0 F 28.4

Min 271.7 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 245.6 F 0.0
10:00 215.2 F 0.0
15:00 236.5 F 0.0
20:00 230.4 F 0.0
25:00 226.0 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 2:34
HEATING: 18:34
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

11/15/2022 03:15 PM
ACCEPTED: X
REJECTED:
Operator: U02