
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001169
13:11 13/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:47
STERILIZING 5:06
136.0 C 222kPa 5:06
136.2 C 224kPa 5:36
136.4 C 225kPa 6:06
136.3 C 225kPa 6:36
136.4 C 225kPa 7:06
136.4 C 226kPa 7:36
136.1 C 223kPa 8:06
136.2 C 224kPa 8:36
136.2 C 224kPa 8:36
VENTING 8:36
AIR DRYING 9:33
CYCLE COMPLETE 19:55

Digital Signature #
4A72B17DB757096E
