
Midmark M11 - v1.0.5
Total Cycles: 2627

Name: RITTER2
Sterilizer ID: V2313228
07/10/2023 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.7 F 0.0
2:00 185.8 F 3.6
4:00 208.8 F 8.9
6:00 222.4 F 8.5
8:00 232.4 F 10.4
10:00 232.7 F 8.1
12:00 240.5 F 11.0
14:00 246.5 F 14.0
16:00 259.0 F 20.8
18:00 270.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.7 F 29.7
2:00 271.7 F 28.9
3:00 271.9 F 29.5
4:00 272.4 F 29.4

Min 271.3 F 28.7
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 229.7 F 0.0
10:00 219.6 F 0.0
15:00 215.1 F 0.0
20:00 210.8 F 0.0
25:00 210.8 F 0.0
30:00 210.8 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 18:54
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

07/10/2023 12:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____