

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1812

Name: RITTER 1  
Sterilizer ID: V2178198  
04/11/2022 02:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
2:00 176.6 F 3.6  
4:00 191.4 F 8.0  
6:00 209.3 F 13.3  
8:00 224.2 F 7.1  
10:00 235.5 F 12.3  
12:00 236.1 F 9.3  
14:00 238.2 F 9.4  
16:00 248.0 F 14.2  
18:00 260.3 F 21.0  
20:00 271.4 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.5 F 28.7  
2:00 271.8 F 28.8  
3:00 272.1 F 28.5  
4:00 272.1 F 29.0  
  
Min 271.2 F 28.2  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.3 F 0.0  
5:00 227.2 F 0.0  
10:00 210.4 F 0.0  
15:00 199.3 F 0.0  
20:00 197.5 F 0.0  
25:00 215.6 F 0.0  
30:00 235.7 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 20:36  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:02  
CYCLE COMPLETE

04/12/2022 08:27 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02