

-----  
Midmark M11 - v1.0.0  
Total Cycles: 835

Name: RITTER 1  
Sterilizer ID: V2178198  
05/26/2021 08:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.8 F 0.0  
2:00 137.5 F 0.5  
4:00 149.8 F 2.5  
6:00 174.7 F 5.4  
8:00 195.6 F 9.1  
10:00 210.9 F 13.3  
12:00 220.1 F 5.8  
14:00 231.2 F 10.0  
16:00 234.7 F 9.6  
18:00 241.0 F 11.6  
20:00 238.5 F 9.8  
22:00 250.9 F 15.8  
24:00 261.6 F 21.9  
26:00 271.5 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.6 F 29.1  
2:00 272.4 F 28.7  
3:00 271.5 F 28.2  
4:00 271.9 F 28.8

Min 271.4 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
5:00 225.8 F 0.0  
10:00 211.7 F 0.0  
15:00 209.3 F 0.0  
20:00 208.6 F 0.0  
25:00 222.9 F 0.0  
30:00 231.9 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 26:43  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:05:07  
CYCLE COMPLETE

05/26/2021 09:41 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_