
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001057
16:18 04/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:47
STERILIZING 5:16
136.4 C 224kPa 5:16
136.3 C 224kPa 5:46
136.3 C 225kPa 6:16
136.4 C 225kPa 6:46
136.4 C 225kPa 7:16
136.4 C 225kPa 7:46
136.4 C 225kPa 8:16
136.2 C 224kPa 8:46
136.2 C 224kPa 8:46
VENTING 8:46
AIR DRYING 9:44
CYCLE COMPLETE 24:45

Digital Signature #
BA2DF0A1EC907516
