
Midmark M11 - v1.0.5
Total Cycles: 1941

Name: RITTER2
Sterilizer ID: V2313228
01/17/2023 04:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.9 F 0.0
2:00 166.2 F 3.1
4:00 180.6 F 7.6
6:00 207.3 F 13.2
8:00 216.8 F 9.0
10:00 234.8 F 9.5
12:00 242.5 F 13.5
14:00 243.6 F 12.9
16:00 257.3 F 20.1
18:00 268.9 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 30.4
1:00 272.3 F 30.5
2:00 272.5 F 30.7
3:00 272.1 F 30.5
4:00 272.5 F 30.4

Min 271.8 F 30.3
Max 273.5 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.8 F 0.0
5:00 240.4 F 0.0
10:00 231.8 F 0.0
15:00 230.7 F 0.0
20:00 233.2 F 0.0
25:00 230.4 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

01/17/2023 05:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____