

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3092

Name: RITTER2  
Sterilizer ID: V2313228  
11/07/2023 10:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.9 F 0.0  
2:00 144.5 F 1.0  
4:00 152.9 F 3.8  
6:00 192.2 F 7.9  
8:00 213.8 F 12.0  
10:00 221.0 F 6.4  
12:00 230.9 F 11.1  
14:00 231.6 F 7.6  
16:00 242.0 F 13.3  
18:00 241.8 F 12.0  
20:00 254.1 F 18.3  
22:00 265.0 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.3  
1:00 272.9 F 30.3  
2:00 271.9 F 30.1  
3:00 271.7 F 29.6  
4:00 272.6 F 30.0

Min 271.0 F 29.1  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.0 F 0.0  
5:00 206.1 F 0.0  
10:00 218.1 F 0.0  
15:00 221.7 F 0.0  
20:00 220.8 F 0.0  
25:00 223.7 F 0.0  
30:00 223.1 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 23:53  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:02:29  
CYCLE COMPLETE

11/07/2023 11:39 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_