
Midmark M11 - v1.0.0
Total Cycles: 2834

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2023 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.9 F 0.0
2:00 206.7 F 4.5
4:00 213.9 F 9.7
6:00 228.3 F 6.2
8:00 234.8 F 10.6
10:00 236.7 F 7.7
12:00 241.9 F 11.1
14:00 246.8 F 13.4
16:00 259.6 F 20.4
18:00 271.1 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.4 F 28.7
2:00 271.5 F 28.1
3:00 272.5 F 28.8
4:00 271.8 F 28.2

Min 271.4 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.8 F 0.0
5:00 228.9 F 0.0
10:00 238.9 F 0.0
15:00 234.6 F 0.0
20:00 226.8 F 0.0
25:00 224.3 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 18:41
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

01/31/2023 02:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____