
Midmark M11 - v1.0.0
Total Cycles: 4759

Name: RITTER 1
Sterilizer ID: V2178198
07/23/2024 10:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.8 F 0.0
2:00 130.7 F 1.2
4:00 140.8 F 3.9
6:00 165.9 F 7.1
8:00 188.9 F 10.3
10:00 204.3 F 5.8
12:00 219.5 F 6.1
14:00 228.1 F 10.1
16:00 234.5 F 8.6
18:00 238.4 F 10.8
20:00 237.6 F 8.3
22:00 248.6 F 14.5
24:00 259.6 F 20.7
26:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.0
2:00 271.6 F 28.2
3:00 272.6 F 29.1
4:00 272.2 F 28.6

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 206.0 F 0.0
10:00 215.3 F 0.0
15:00 223.1 F 0.0
20:00 226.0 F 0.0
25:00 230.0 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:59
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:35
CYCLE COMPLETE

07/23/2024 11:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____