
Midmark M11 - v1.0.0
Total Cycles: 1580

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2022 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.6 F 0.0
2:00 180.3 F 2.8
4:00 185.9 F 7.1
6:00 211.8 F 13.0
8:00 224.2 F 7.3
10:00 237.3 F 12.6
12:00 237.4 F 9.8
14:00 238.5 F 8.5
16:00 249.9 F 14.8
18:00 261.9 F 21.7
20:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.6 F 28.7
2:00 271.4 F 27.9
3:00 272.7 F 28.9
4:00 272.1 F 28.3

Min 271.3 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 230.4 F 0.0
10:00 232.6 F 0.0
15:00 231.1 F 0.0
20:00 230.0 F 0.0
25:00 227.6 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:12
CYCLE COMPLETE

02/01/2022 11:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____