
Midmark M11 - v1.0.0
Total Cycles: 4929

Name: RITTER 1
Sterilizer ID: V2178198
09/23/2024 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 71.1 F 0.0
4:00 104.1 F 0.0
6:00 116.7 F 0.0
8:00 144.6 F 0.2
10:00 173.1 F 0.4
12:00 193.3 F 1.0
14:00 210.8 F 4.5
16:00 218.9 F 7.8
18:00 229.1 F 11.4
20:00 227.8 F 6.0
22:00 239.9 F 11.6
24:00 236.5 F 9.0
26:00 239.6 F 9.7
28:00 248.4 F 14.2
30:00 260.3 F 20.8
32:00 270.9 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 27.9
1:00 271.9 F 28.3
2:00 272.6 F 28.6
3:00 271.9 F 28.0
4:00 271.6 F 28.1

Min 271.4 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.3 F 0.0
5:00 192.5 F 0.0
10:00 197.9 F 0.0
15:00 209.2 F 0.0
20:00 211.6 F 0.0
25:00 213.3 F 0.0
30:00 214.4 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:41
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:10:50
CYCLE COMPLETE

09/23/2024 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____