
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002067
9:34 20/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:31
PRESSURIZING 1:38
STERILIZING 2:35
136.0 C 221kPa 2:35
136.2 C 225kPa 3:05
136.4 C 226kPa 3:35
136.3 C 225kPa 4:05
136.4 C 224kPa 4:35
136.4 C 226kPa 5:05
136.4 C 226kPa 5:35
136.3 C 223kPa 6:05
136.3 C 223kPa 6:05
VENTING 6:05
AIR DRYING 6:55
CYCLE COMPLETE 13:35

Digital Signature #
B31DC523E24896B9
