
Midmark M11 - v1.0.0
Total Cycles: 2063

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2022 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.3 F 0.0
2:00 140.0 F 0.8
4:00 162.5 F 3.6
6:00 185.5 F 7.6
8:00 219.0 F 12.7
10:00 223.5 F 6.8
12:00 236.3 F 11.7
14:00 234.5 F 8.5
16:00 242.1 F 10.6
18:00 246.0 F 13.2
20:00 258.4 F 19.9
22:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 271.8 F 28.9
2:00 272.2 F 28.8
3:00 271.1 F 28.3
4:00 272.4 F 29.2

Min 271.1 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.6 F 0.0
5:00 218.9 F 0.0
10:00 209.8 F 0.0
15:00 208.4 F 0.0
20:00 246.4 F 0.0
25:00 242.9 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 22:49
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:56
CYCLE COMPLETE

06/27/2022 04:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____