
Midmark M11 - v1.0.0
Total Cycles: 3497

Name: RITTER 1
Sterilizer ID: V2178198
07/06/2023 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.8 F 0.0
2:00 186.7 F 2.8
4:00 201.2 F 7.5
6:00 218.9 F 13.0
8:00 226.9 F 8.8
10:00 236.7 F 10.2
12:00 243.2 F 12.4
14:00 241.9 F 11.3
16:00 255.5 F 18.3
18:00 267.4 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.3 F 29.0
2:00 271.4 F 28.6
3:00 272.3 F 29.0
4:00 271.3 F 28.4

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 213.2 F 0.0
10:00 223.6 F 0.0
15:00 229.8 F 0.0
20:00 228.0 F 0.0
25:00 228.3 F 0.0
30:00 230.7 F 0.0

Drying Complete

FILLING: 1:38
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

07/06/2023 02:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____