
Midmark M11 - v1.0.5
Total Cycles: 2193

Name: RITTER2
Sterilizer ID: V2313228
03/29/2023 11:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.3 F 0.0
2:00 165.8 F 3.1
4:00 177.8 F 7.0
6:00 197.5 F 12.1
8:00 217.4 F 6.7
10:00 231.4 F 12.0
12:00 234.1 F 9.4
14:00 237.7 F 9.5
16:00 248.2 F 15.0
18:00 260.6 F 21.9
20:00 271.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.7 F 29.9
2:00 271.7 F 29.2
3:00 272.0 F 29.7
4:00 272.6 F 29.7

Min 271.2 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 204.9 F 0.0
10:00 210.3 F 0.0
15:00 216.4 F 0.0
20:00 212.8 F 0.0
25:00 216.0 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:59:12
CYCLE COMPLETE

03/29/2023 12:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____