
Midmark M11 - v1.0.5
Total Cycles: 89

Name: RITTER2
Sterilizer ID: V2313228
06/10/2021 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.9 F 0.0
2:00 158.8 F 2.9
4:00 177.1 F 6.7
6:00 197.5 F 11.1
8:00 213.3 F 5.0
10:00 224.3 F 9.9
12:00 232.5 F 9.2
14:00 239.6 F 12.8
16:00 241.8 F 11.9
18:00 254.0 F 18.1
20:00 264.4 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 271.9 F 29.5
2:00 271.5 F 29.6
3:00 271.9 F 29.4
4:00 271.6 F 29.7

Min 271.2 F 29.0
Max 272.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 219.6 F 0.0
10:00 221.4 F 0.0
15:00 216.7 F 0.0
20:00 213.5 F 0.0
25:00 210.7 F 0.0
30:00 207.7 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 21:56
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:23
CYCLE COMPLETE

06/10/2021 02:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____