
Midmark M11 - v1.0.0
Total Cycles: 2308

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2022 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.1 F 0.0
2:00 150.3 F 2.4
4:00 172.1 F 5.6
6:00 193.5 F 9.6
8:00 210.8 F 7.4
10:00 222.7 F 7.7
12:00 232.4 F 12.6
14:00 234.3 F 9.0
16:00 240.6 F 10.3
18:00 244.7 F 12.8
20:00 256.8 F 19.2
22:00 267.6 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.2
2:00 271.2 F 28.5
3:00 272.4 F 29.2
4:00 271.3 F 28.4

Min 271.2 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 241.3 F 0.0
10:00 227.3 F 0.0
15:00 229.4 F 0.0
20:00 224.2 F 0.0
25:00 229.0 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:23
CYCLE COMPLETE

09/12/2022 10:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____