
Midmark M11 - v1.0.5
Total Cycles: 3260

Name: RITTER2
Sterilizer ID: V2313228
12/28/2023 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.6 F 0.0
2:00 144.8 F 1.1
4:00 156.2 F 3.8
6:00 179.7 F 8.0
8:00 204.8 F 13.4
10:00 224.5 F 7.7
12:00 236.2 F 12.8
14:00 235.9 F 9.7
16:00 236.6 F 8.9
18:00 247.9 F 14.7
20:00 260.1 F 21.5
22:00 271.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 273.0 F 29.7
2:00 272.4 F 30.0
3:00 272.5 F 29.4
4:00 272.6 F 30.1

Min 271.5 F 29.2
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.4 F 0.0
5:00 244.1 F 0.0
10:00 210.1 F 0.0
15:00 208.3 F 0.0
20:00 184.0 F 0.0
25:00 181.4 F 0.0
30:00 180.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:31
CYCLE COMPLETE

12/28/2023 12:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____