

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3860

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2023 11:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.5 F 0.0  
2:00 167.8 F 2.9  
4:00 179.9 F 6.4  
6:00 200.7 F 10.8  
8:00 221.2 F 5.1  
10:00 227.6 F 9.5  
12:00 236.2 F 10.2  
14:00 239.9 F 11.1  
16:00 237.0 F 9.1  
18:00 250.4 F 15.4  
20:00 261.3 F 21.6  
22:00 271.3 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.0 F 28.9  
2:00 272.4 F 29.0  
3:00 272.2 F 28.9  
4:00 272.1 F 28.9

Min 271.4 F 28.4  
Max 272.4 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.8 F 0.0  
5:00 228.9 F 0.0  
10:00 230.2 F 0.0  
15:00 230.1 F 0.0  
20:00 226.1 F 0.0  
25:00 214.6 F 0.0  
30:00 202.8 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 22:31  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:01:49  
CYCLE COMPLETE

10/04/2023 12:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_