

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2910

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2023 12:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.4 F 0.0  
2:00 176.7 F 2.8  
4:00 191.1 F 7.3  
6:00 207.5 F 12.2  
8:00 219.9 F 7.1  
10:00 233.7 F 12.3  
12:00 235.6 F 10.0  
14:00 236.5 F 9.0  
16:00 248.8 F 15.3  
18:00 260.3 F 21.8  
20:00 270.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.6 F 29.9  
2:00 271.2 F 29.2  
3:00 272.7 F 30.0  
4:00 271.7 F 29.4  
  
Min 271.1 F 29.2  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.9 F 0.0  
5:00 211.4 F 0.0  
10:00 223.4 F 0.0  
15:00 217.5 F 0.0  
20:00 214.6 F 0.0  
25:00 215.4 F 0.0  
30:00 215.1 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 20:47  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:00:09  
CYCLE COMPLETE

09/20/2023 01:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_