
Midmark M11 - v1.0.0
Total Cycles: 4997

Name: RITTER 1
Sterilizer ID: V2178198
10/16/2024 08:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 76.6 F 0.0
4:00 93.3 F 0.0
6:00 130.8 F 0.1
8:00 152.4 F 0.2
10:00 180.2 F 1.0
12:00 196.1 F 4.2
14:00 206.2 F 7.3
16:00 218.2 F 10.5
18:00 223.5 F 8.5
20:00 231.0 F 8.8
22:00 241.8 F 13.4
24:00 240.3 F 10.8
26:00 236.4 F 8.7
28:00 250.7 F 15.4
30:00 262.1 F 21.8
32:00 272.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.4 F 28.4
2:00 271.9 F 28.3
3:00 272.3 F 28.3
4:00 271.9 F 28.4

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 191.5 F 0.0
10:00 190.9 F 0.0
15:00 190.5 F 0.0
20:00 190.4 F 0.0
25:00 205.1 F 0.0
30:00 210.8 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:24
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:10:31
CYCLE COMPLETE

10/16/2024 09:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____