

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002524  
11:55 20/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:46  
STERILIZING 3:54  
136.4 C 223kPa 3:54  
136.1 C 222kPa 4:24  
136.4 C 223kPa 4:54  
136.4 C 225kPa 5:24  
136.3 C 224kPa 5:54  
136.3 C 224kPa 6:24  
136.4 C 225kPa 6:54  
136.4 C 225kPa 7:24  
136.4 C 225kPa 7:24  
VENTING 7:24  
AIR DRYING 8:16  
CYCLE COMPLETE 8:43

-----  
Digital Signature #  
51BB9D05C0488D8D  
-----