
Midmark M11 - v1.0.0
Total Cycles: 1669

Name: RITTER 1
Sterilizer ID: V2178198
02/24/2022 12:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.0 F 0.0
2:00 167.3 F 1.8
4:00 184.3 F 5.8
6:00 206.3 F 10.9
8:00 219.4 F 4.6
10:00 232.3 F 9.8
12:00 235.8 F 8.7
14:00 243.1 F 12.1
16:00 240.6 F 10.3
18:00 253.9 F 17.0
20:00 265.6 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.2 F 28.4
2:00 271.5 F 28.3
3:00 272.2 F 28.4
4:00 271.4 F 28.2

Min 271.2 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 225.4 F 0.0
10:00 219.9 F 0.0
15:00 224.2 F 0.0
20:00 218.0 F 0.0
25:00 200.7 F 0.0
30:00 195.1 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:44
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

02/24/2022 01:39 PM
ACCEPTED: X
REJECTED:
Operator: U02