

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1112

Name: RITTER 1  
Sterilizer ID: V2178198  
08/24/2021 06:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.5 F 0.0  
2:00 78.4 F 0.0  
4:00 116.3 F 0.0  
6:00 116.5 F 0.0  
8:00 143.2 F 0.1  
10:00 175.2 F 0.2  
12:00 197.9 F 1.4  
14:00 215.3 F 4.5  
16:00 222.0 F 7.0  
18:00 231.6 F 10.4  
20:00 233.9 F 9.0  
22:00 240.1 F 11.1  
24:00 235.5 F 8.6  
26:00 240.8 F 10.3  
28:00 247.7 F 13.8  
30:00 259.6 F 20.3  
32:00 270.5 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.7 F 28.7  
2:00 271.5 F 27.9  
3:00 272.8 F 28.8  
4:00 271.9 F 28.1  
  
Min 271.5 F 27.8  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.4 F 0.0  
5:00 208.8 F 0.0  
10:00 203.3 F 0.0  
15:00 197.8 F 0.0  
20:00 208.3 F 0.0  
25:00 219.1 F 0.0  
30:00 231.3 F 0.0  
  
Drying Complete

FILLING: 1:25  
HEATING: 32:51  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:11:20  
CYCLE COMPLETE

08/24/2021 08:42 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02