

-----  
Midmark M11 - v1.0.0  
Total Cycles: 976

Name: RITTER 1  
Sterilizer ID: V2178198  
07/07/2021 01:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 114.1 F 0.0  
2:00 118.0 F 0.1  
4:00 145.4 F 0.9  
6:00 174.9 F 3.9  
8:00 192.6 F 7.6  
10:00 212.7 F 12.1  
12:00 219.8 F 5.7  
14:00 232.3 F 10.4  
16:00 233.5 F 8.1  
18:00 243.4 F 12.6  
20:00 241.4 F 11.0  
22:00 254.0 F 17.3  
24:00 265.1 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.5 F 28.8  
2:00 271.5 F 28.1  
3:00 272.6 F 28.9  
4:00 271.7 F 28.2

Min 271.5 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
5:00 205.4 F 0.0  
10:00 216.5 F 0.0  
15:00 217.5 F 0.0  
20:00 220.7 F 0.0  
25:00 225.9 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 25:52  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:03:59  
CYCLE COMPLETE

07/07/2021 02:08 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01