
Midmark M11 - v1.0.0
Total Cycles: 738

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2021 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.3 F 0.0
2:00 123.4 F 0.2
4:00 129.0 F 2.7
6:00 161.7 F 6.2
8:00 230.7 F 11.1
10:00 227.0 F 5.8
12:00 238.6 F 11.2
14:00 235.0 F 8.0
16:00 240.9 F 10.6
18:00 245.9 F 12.8
20:00 258.5 F 19.5
22:00 269.6 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.4 F 28.9
2:00 272.2 F 28.6
3:00 271.5 F 28.4
4:00 271.6 F 28.8

Min 271.3 F 28.4
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 242.6 F 0.0
10:00 222.9 F 0.0
15:00 221.8 F 0.0
20:00 223.8 F 0.0
25:00 229.5 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:31
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:01:23
CYCLE COMPLETE

03/22/2021 03:55 PM
ACCEPTED: X
REJECTED:
Operator: U01