
Midmark M11 - v1.0.0
Total Cycles: 3447

Name: RITTER 1
Sterilizer ID: V2178198
06/26/2023 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.7 F 0.0
2:00 198.8 F 3.6
4:00 205.7 F 8.2
6:00 221.8 F 10.6
8:00 228.8 F 8.8
10:00 240.4 F 11.1
12:00 241.8 F 12.0
14:00 240.1 F 10.6
16:00 253.5 F 17.2
18:00 265.1 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.8 F 29.2
2:00 271.7 F 28.4
3:00 272.6 F 29.3
4:00 272.3 F 28.9

Min 271.5 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 228.4 F 0.0
10:00 233.0 F 0.0
15:00 229.1 F 0.0
20:00 228.4 F 0.0
25:00 229.3 F 0.0
30:00 229.9 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:12
CYCLE COMPLETE

06/26/2023 03:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____