

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4512

Name: RITTER 1  
Sterilizer ID: V2178198  
05/01/2024 03:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.3 F 0.0  
2:00 167.4 F 3.7  
4:00 185.8 F 8.1  
6:00 205.5 F 13.2  
8:00 223.3 F 7.5  
10:00 235.4 F 12.8  
12:00 237.3 F 10.2  
14:00 236.0 F 8.3  
16:00 249.6 F 15.2  
18:00 261.5 F 21.9  
20:00 272.2 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.4 F 29.0  
2:00 271.5 F 28.8  
3:00 272.5 F 29.0  
4:00 271.3 F 28.5  
  
Min 271.2 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.1 F 0.0  
5:00 218.9 F 0.0  
10:00 230.7 F 0.0  
15:00 230.0 F 0.0  
20:00 229.6 F 0.0  
25:00 229.7 F 0.0  
30:00 228.4 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 20:29  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 00:59:45  
CYCLE COMPLETE

05/01/2024 04:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_