
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000949
17:19 23/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:46
STERILIZING 4:14
136.4 C 232kPa 4:14
136.1 C 231kPa 4:44
136.4 C 233kPa 5:14
136.4 C 230kPa 5:44
136.1 C 231kPa 6:14
136.2 C 231kPa 6:44
136.3 C 229kPa 7:14
136.3 C 232kPa 7:44
136.3 C 232kPa 7:44
VENTING 7:44
AIR DRYING 8:37
CYCLE COMPLETE 17:22

Digital Signature #
DDDD2DFB002CD5BB
