
Midmark M11 - v1.0.5
Total Cycles: 3498

Name: RITTER2
Sterilizer ID: V2313228
04/05/2024 08:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.0 F 0.0
2:00 68.2 F 0.0
4:00 74.6 F 0.0
6:00 94.3 F 0.1
8:00 149.2 F 0.3
10:00 189.5 F 2.3
12:00 197.0 F 6.0
14:00 209.1 F 9.4
16:00 232.2 F 12.9
18:00 226.8 F 7.3
20:00 235.5 F 11.7
22:00 233.7 F 8.6
24:00 241.8 F 11.7
26:00 244.4 F 12.9
28:00 256.2 F 19.1
30:00 266.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.5 F 29.7
2:00 271.3 F 28.8
3:00 272.7 F 29.5
4:00 271.2 F 28.8

Min 271.1 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 233.3 F 0.0
10:00 222.8 F 0.0
15:00 176.5 F 0.0
20:00 173.1 F 0.0
25:00 171.7 F 0.0
30:00 171.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:22
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:32
CYCLE COMPLETE

04/05/2024 09:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____