

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3514

Name: RITTER2  
Sterilizer ID: V2313228  
04/10/2024 03:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 122.0 F 0.0  
2:00 122.4 F 0.7  
4:00 128.4 F 3.2  
6:00 155.3 F 6.3  
8:00 175.7 F 10.0  
10:00 203.4 F 8.2  
12:00 216.4 F 8.0  
14:00 230.8 F 12.4  
16:00 230.9 F 8.8  
18:00 239.2 F 11.4  
20:00 242.6 F 12.6  
22:00 254.3 F 18.6  
24:00 264.9 F 25.0  
26:00 271.0 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.5  
1:00 272.3 F 30.5  
2:00 272.6 F 30.2  
3:00 272.2 F 30.7  
4:00 272.6 F 30.2

Min 271.0 F 29.3  
Max 273.5 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.9 F 0.0  
5:00 210.4 F 0.0  
10:00 222.1 F 0.0  
15:00 227.3 F 0.0  
20:00 228.3 F 0.0  
25:00 227.0 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 26:01  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:04:37  
CYCLE COMPLETE

04/10/2024 04:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_