
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008209
8:59 15/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:12
STERILIZING 4:18
136.4 C 223kPa 4:18
136.3 C 223kPa 4:48
136.4 C 220kPa 5:18
136.4 C 219kPa 5:48
136.5 C 223kPa 6:18
136.4 C 223kPa 6:48
136.4 C 221kPa 7:18
136.1 C 219kPa 7:48
136.1 C 219kPa 7:48
VENTING 7:48
AIR DRYING 8:48
CYCLE COMPLETE 38:48

Digital Signature #
05936834BDC862AE
