
Midmark M11 - v1.0.0
Total Cycles: 2389

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2022 01:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.9 F 0.0
2:00 196.9 F 3.6
4:00 204.5 F 8.1
6:00 223.3 F 9.1
8:00 231.8 F 9.2
10:00 239.3 F 9.4
12:00 243.8 F 12.6
14:00 242.8 F 11.5
16:00 256.1 F 18.4
18:00 267.6 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.7
2:00 272.1 F 28.8
3:00 272.1 F 28.5
4:00 272.2 F 29.0

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 243.0 F 0.0
10:00 233.6 F 0.0
15:00 231.3 F 0.0
20:00 227.3 F 0.0
25:00 228.3 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:47
CYCLE COMPLETE

10/03/2022 02:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____