

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3507

Name: RITTER 1  
Sterilizer ID: V2178198  
07/10/2023 12:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
2:00 188.0 F 3.6  
4:00 194.7 F 7.9  
6:00 219.3 F 13.0  
8:00 226.1 F 7.9  
10:00 237.9 F 13.0  
12:00 237.9 F 10.4  
14:00 235.0 F 8.4  
16:00 249.9 F 15.3  
18:00 261.6 F 22.0  
20:00 272.7 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.5  
1:00 272.7 F 29.2  
2:00 271.6 F 28.6  
3:00 272.4 F 28.9  
4:00 271.8 F 28.9  
  
Min 271.5 F 28.4  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
5:00 217.3 F 0.0  
10:00 235.7 F 0.0  
15:00 228.9 F 0.0  
20:00 226.3 F 0.0  
25:00 222.6 F 0.0  
30:00 217.9 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 20:14  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:07  
CYCLE COMPLETE

07/10/2023 01:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_