
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000318
10:29 04/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 3:24
STERILIZING 5:00
136.4 C 232kPa 5:00
136.4 C 233kPa 5:30
136.4 C 233kPa 6:00
136.4 C 233kPa 6:30
136.3 C 232kPa 7:00
136.5 C 233kPa 7:30
136.4 C 233kPa 8:00
136.4 C 233kPa 8:30
136.4 C 233kPa 8:30
VENTING 8:30
CYCLE ABORTED 9:12
STOP BUTTON PRESSED
NOT STERILE
END TIME 9:12

Digital Signature #
827740F76094616A
