
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006637
9:38 17/11/2023

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:24
STERILIZING 3:20
136.0 C 220kPa 3:20
136.4 C 221kPa 3:50
136.4 C 219kPa 4:20
136.4 C 222kPa 4:50
136.4 C 220kPa 5:20
136.4 C 221kPa 5:50
136.1 C 221kPa 6:20
136.3 C 222kPa 6:50
136.3 C 222kPa 6:50
VENTING 6:50
AIR DRYING 7:43
CYCLE COMPLETE 37:43

Digital Signature #
21273B6A329D0A16
