
Midmark M11 - v1.0.5
Total Cycles: 2647

Name: RITTER2
Sterilizer ID: V2313228
07/14/2023 11:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.0 F 0.0
2:00 180.1 F 1.9
4:00 181.8 F 5.7
6:00 202.4 F 10.5
8:00 215.3 F 5.4
10:00 229.3 F 9.2
12:00 236.5 F 11.1
14:00 241.6 F 12.0
16:00 240.6 F 10.9
18:00 253.3 F 17.4
20:00 264.2 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.0 F 29.5
2:00 272.3 F 29.2
3:00 271.4 F 29.0
4:00 272.5 F 29.7

Min 271.3 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 217.3 F 0.0
10:00 212.3 F 0.0
15:00 210.0 F 0.0
20:00 206.2 F 0.0
25:00 203.3 F 0.0
30:00 202.9 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 21:57
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:60
CYCLE COMPLETE

07/14/2023 12:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____