

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000261  
14:21 30/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:35  
PRESSURIZING 3:35  
STERILIZING 5:02  
136.4 C 233kPa 5:02  
136.4 C 233kPa 5:32  
136.4 C 233kPa 6:02  
136.3 C 233kPa 6:32  
136.4 C 233kPa 7:02  
136.4 C 233kPa 7:32  
136.3 C 233kPa 8:02  
136.4 C 233kPa 8:32  
136.4 C 233kPa 8:32  
VENTING 8:32  
AIR DRYING 9:34  
CYCLE COMPLETE 24:34  
-----

Digital Signature #  
6218ED5D81726255  
-----