

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5116

Name: RITTER 1  
Sterilizer ID: V2178198  
11/20/2024 04:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.5 F 0.0  
2:00 124.6 F 0.0  
4:00 146.3 F 0.2  
6:00 178.2 F 0.6  
8:00 199.8 F 1.6  
10:00 218.7 F 5.4  
12:00 226.7 F 9.0  
14:00 237.2 F 13.0  
16:00 236.5 F 9.2  
18:00 241.1 F 10.9  
20:00 244.8 F 12.5  
22:00 242.5 F 11.4  
24:00 254.8 F 17.6  
26:00 265.5 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.6 F 28.9  
2:00 272.0 F 28.3  
3:00 271.4 F 28.1  
4:00 272.4 F 28.9

Min 271.3 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 173.3 F 0.0  
5:00 234.0 F 0.0  
10:00 196.7 F 0.0  
15:00 196.1 F 0.0  
20:00 196.6 F 0.0  
25:00 192.9 F 0.0  
30:00 211.7 F 0.0  
Drying Complete

FILLING: 0:40  
HEATING: 27:52  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:06:21  
CYCLE COMPLETE

11/20/2024 05:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_