
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002178
9:50 29/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:37
STERILIZING 3:37
136.0 C 220kPa 3:37
136.1 C 220kPa 4:07
136.2 C 223kPa 4:37
136.3 C 223kPa 5:07
136.3 C 224kPa 5:37
136.2 C 223kPa 6:07
136.2 C 222kPa 6:37
136.4 C 225kPa 7:07
136.4 C 225kPa 7:07
VENTING 7:07
AIR DRYING 7:57
CYCLE COMPLETE 17:34

Digital Signature #
50D58D601E9AE646
