
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006973
11:29 23/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:58
STERILIZING 2:53
136.4 C 222kPa 2:53
136.3 C 224kPa 3:23
136.1 C 217kPa 3:53
136.2 C 222kPa 4:23
136.4 C 224kPa 4:53
136.1 C 221kPa 5:23
136.3 C 223kPa 5:53
136.4 C 222kPa 6:23
136.4 C 222kPa 6:23
VENTING 6:23
AIR DRYING 7:16
CYCLE COMPLETE 37:16

Digital Signature #
4FE9B49A8CF149E0
