

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4612

Name: RITTER 1  
Sterilizer ID: V2178198  
06/06/2024 08:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 73.7 F 0.2  
4:00 100.4 F 0.4  
6:00 125.1 F 1.1  
8:00 145.8 F 3.5  
10:00 172.4 F 6.4  
12:00 190.1 F 10.0  
14:00 203.0 F 7.9  
16:00 213.0 F 5.5  
18:00 224.4 F 9.1  
20:00 234.9 F 13.2  
22:00 234.3 F 9.3  
24:00 239.7 F 10.3  
26:00 244.5 F 12.8  
28:00 256.2 F 18.9  
30:00 266.6 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.7 F 29.3  
2:00 271.6 F 28.5  
3:00 272.7 F 29.4  
4:00 272.1 F 28.8

Min 271.5 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
5:00 198.3 F 0.0  
10:00 205.3 F 0.0  
15:00 213.0 F 0.0  
20:00 217.6 F 0.0  
25:00 222.8 F 0.0  
30:00 225.1 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:38  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:09:51  
CYCLE COMPLETE

06/06/2024 09:22 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_