
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004700
11:39 14/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:29
STERILIZING 3:17
136.0 C 220kPa 3:17
136.3 C 224kPa 3:47
136.4 C 225kPa 4:17
136.3 C 221kPa 4:47
136.4 C 225kPa 5:17
136.3 C 224kPa 5:47
136.4 C 224kPa 6:17
136.4 C 224kPa 6:47
136.4 C 224kPa 6:47
VENTING 6:47
AIR DRYING 7:36
CYCLE COMPLETE 10:10

Digital Signature #
67DB394F8D8CAF66
