

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3674

Name: RITTER2  
Sterilizer ID: V2313228  
06/06/2024 02:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.0 F 0.0  
2:00 97.7 F 0.5  
4:00 113.1 F 2.4  
6:00 146.1 F 5.2  
8:00 174.0 F 8.8  
10:00 200.0 F 12.8  
12:00 211.2 F 6.1  
14:00 220.4 F 10.4  
16:00 228.4 F 8.1  
18:00 237.6 F 11.4  
20:00 235.6 F 9.4  
22:00 248.9 F 15.6  
24:00 260.2 F 21.9  
26:00 270.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.6 F 29.9  
1:00 272.9 F 30.6  
2:00 272.2 F 29.8  
3:00 273.1 F 30.6  
4:00 272.4 F 29.8

Min 271.9 F 29.7  
Max 273.1 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.6 F 0.0  
5:00 229.7 F 0.0  
10:00 210.7 F 0.0  
15:00 184.9 F 0.0  
20:00 178.7 F 0.0  
25:00 178.8 F 0.0  
30:00 193.6 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 27:27  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:05:42  
CYCLE COMPLETE

06/06/2024 03:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_