
Midmark M11 - v1.0.5
Total Cycles: 1526

Name: RITTER2
Sterilizer ID: V2313228
09/22/2022 01:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 185.9 F 3.6
4:00 188.0 F 8.3
6:00 207.4 F 12.0
8:00 224.4 F 8.9
10:00 234.9 F 10.2
12:00 240.7 F 12.2
14:00 240.3 F 11.2
16:00 253.2 F 17.8
18:00 264.8 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.2 F 30.1
2:00 272.3 F 29.7
3:00 271.3 F 29.2
4:00 271.5 F 29.7

Min 271.1 F 29.1
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 209.6 F 0.0
10:00 219.6 F 0.0
15:00 218.4 F 0.0
20:00 223.3 F 0.0
25:00 232.2 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 19:58
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

09/22/2022 02:37 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____