
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001764
16:35 01/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:54
STERILIZING 2:55
136.5 C 226kPa 2:55
136.2 C 224kPa 3:25
136.4 C 226kPa 3:55
136.3 C 225kPa 4:25
136.4 C 225kPa 4:55
136.4 C 226kPa 5:25
136.4 C 223kPa 5:55
136.0 C 222kPa 6:25
136.0 C 222kPa 6:25
VENTING 6:25
AIR DRYING 7:14
CYCLE COMPLETE 7:22

Digital Signature #
2428067BA20710DA
