
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 005145
13:12 21/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 1:59
STERILIZING 2:48
136.0 C 220kPa 2:48
136.5 C 225kPa 3:18
136.4 C 223kPa 3:48
136.3 C 225kPa 4:18
136.4 C 224kPa 4:48
136.4 C 223kPa 5:18
136.4 C 223kPa 5:48
136.4 C 222kPa 6:18
136.4 C 222kPa 6:18
VENTING 6:18
AIR DRYING 7:08
CYCLE COMPLETE 7:14

Digital Signature #
548D7141DB6989FF
